



### How to use keypad



### Battery pack




- Connect battery pack (LBT-20) with AC charger, power cable.
- Do not stack battery pack on top of charger while recharging.
- Recharging temperature: 0~40°C (32~104°F).



Confirm power saving function is working when using battery pack.


### Splicing operation



Insert fiber into the protection sleeve


- Make sure the stripped fiber is free of coating debris or dust.
- Use only 99% or better purity alcohol.
- Clean by clean wipes (dustproof)

- Keep the cleaved fiber ends face in clean condition.
- Place the fiber end between V-groove edge and Electrode center.



- Hold the fiber at edge of splicer body
- Gently pull the fiber

- Turning splicer on
- Select splice mode
  - When splicing only standard SM fibers (ITU-T G.652), "SM AUTO" mode is recommended.
  - When splicing different types of fibers, "AUTO" mode is recommended, but splice speed is slower.
  - Splicing speed of "SM FAST" mode is faster, but periodically Arc calibration is required.
- Check splicing and heater modes
- Cleaning fibers
- Place the sleeve on the fiber
- Stripping fiber
- Cleaning fiber
- Cleaving fiber
- Install fiber onto splicer
- Start splicing by closing wind proof cover
- Visual inspection on LCD during splice
- Remove the spliced fiber
- Centering protection sleeve on spliced fiber
- Put the sleeved fiber into heater
- Start heating by closing heater cover
- Completed



When ARC discharge is not stable. Electrodes should be stabilized and calibrated to reform the arc discharge. Install prepared SM fibers onto splicer to operate stabilize electrodes and arc calibration. After installation SM fibers, splice fibers again.

## Maintenance menu

### Maintenance



Prepare SM fiber and go to "Main Menu"



Execute [1. Quick Optimize]




Execute [9. Arc Calibration]



Execute [4. Stabilize Electrodes]

### ARC Calibration



**Important:** 1) Always use SM fiber only 2) To keep in best splice condition, following the procedure.

### Diagnostic test

In case of trouble, proceed this procedure.

Operation procedure :

- Select [Diagnostic Test] in [Maintenance Menu] and execute [Diagnostic Test].
  - LED Check
 

Measure and adjust the brightness of the illumination LED.
  - Dust Check
 

Check the optical path for dust or dirt and judges whether they disturb fiber observation.

If contamination exists, this function indicates the location.
  - Motor Calibration
 



Automatically calibrate press motor movement.
  - Arc Calibration
 

Automatically calibrate the arc power value.
- Upon completion of all checks and adjustments, a list of results is displayed. If the dust check result is not good, clean the objective lenses. If the cleaning cannot eliminate contamination, there is a possibility that the contamination may be in the inside of the optical path. Please contact your sales agency for additional instructions. The dust check and motor calibration functions exist as independent instructions in [Maintenance Menu]. It is possible to execute them independently.


*Note : Before you start test, remove the fibers from the splicer. When the above item No. 2 (dust check) is completed, prepare and load the fibers onto the splicer and press "SET"*

## Cleaning before splicing

### V-grooves



### Objective lens



When lens is dirty, clean it.


### CAUTION

- During cleaning,
  - Do not contact the electrode tips.
  - Use only 99% or better purity alcohol.
  - Clean by clean wipes (dustproof)

### Fiber clamp chips



### Wind proof cover mirrors



## Periodical maintenance

### Objective lens

If lens is dirty, clean it periodically, with alcohol moistened thin cotton swab. Otherwise, the lens can make trouble.

### Electrodes

When "Replace Electrodes" message appears, or when the tip is damaged, replace electrodes.



- Execute [Replace Electrodes]
- Press "SET", then automatically turns off
- Loosen screws
- Take out electrodes
- Put new electrodes
- Reassemble the electrode cover
- Tighten screws
- Turn splicer on
- Install prepared fibers
- Execute [Stabilize Electrodes]
- Execute [Arc Calibration]

### Battery pack

- Discharge battery pack completely every 3 months before recharging.
- When storing battery pack in long term;
  - Storing temperature should be -20~-30°C (-4~-86°F)
  - Recharge battery pack every 6 months, not to be empty.